



# Certificate of Traceability

## Level I Laboratory Certification

This certifies that Centralab, using Bio-Rad VII Turbo 2.0 has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:

Instrument: <b>VII Turbo 2.0</b>	Calibrator Lot: <b>64450601, 64450602</b>	Column Lot: <b>90072</b>
Reagent Lot: <b>64458630, 64458640</b>	Calibrator Assigned Values: <b>5.60%, 10.30%</b>	

Date of Certification: January 1, 2023      Certification Expires: January 1, 2024

David Searls  
NGSP Steering Committee Chair

Randa Little PhD.  
NGSP Network Coordinator

Shan Amalf  
SRL director/ supervisor