

Certificate of Traceability

Level I Laboratory Certification

This certifies that Centralab, using Bio-Rad Variant II Turbo 2.0 has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: University of Missouri SRL#9

The system evaluated was:


Instrument: Variant II Turbo 2.0	Calibrator Lot: 64487721, 64487722	Column Lot: 90072
Reagent Lot: 64503421, 64494997	Calibrator Assigned Values: 5.30%, 10.50%	

Date of Certification: January 1, 2024

Certification Expires: January 1, 2025


NGSP Steering Committee Chair


NGSP Network Coordinator


SRL director/ supervisor