



Certificate of Traceability

Level I Laboratory Certification

This certifies that Centralab, using Bio-Rad Variant II Turbo 2.0 has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial Reference method**. The comparison was performed with: University of Missouri SRL#12

The system evaluated was:

Instrument: Variant II Turbo 2.0	Calibrator Lot: 64596761, 64596762	Column Lot: 10256
Reagent Lot: 64619064, 64613504	Calibrator Assigned Values: 5.50%, 10.10%	

Date of Certification: January 1, 2025

Certification Expires: January 1, 2026

Daniel South

NGSP Steering Committee Chair

Randa Little PhD.

NGSP Network Coordinator

Shan Amulfi

SRL director/ supervisor