



# Certificate of Traceability

## Level I Laboratory Certification

This certifies that Centralab, using Bio-Rad Variant II Turbo 2.0 has participated in and successfully completed NGSP Level I Laboratory certification and is traceable to the **Diabetes Control and Complications Trial** Reference method. The comparison was performed with: University of Missouri SRL#12

The system evaluated was:

Instrument: <b>Variant II Turbo 2.0</b>	Calibrator Lot: <b>S50101</b>	Column Lot: <b>64683530</b>
Reagent Lot: <b>64683247, 64671886, 64677167</b>	Calibrator Assigned Values: <b>5.3%, 10.3%</b>	

Date of Certification: January 1, 2026

Certification Expires: January 1, 2027

A handwritten signature in black ink, appearing to read 'David Sacks', is written over a horizontal line.

NGSP Steering Committee Chair

A handwritten signature in black ink, appearing to read 'Curt Rohlfing', is written over a horizontal line.

NGSP Network Coordinator

A handwritten signature in black ink, consisting of a stylized monogram, is written over a horizontal line.

SRL director/ supervisor